

# 计量标准考核证书

Certificate for Examination of Measurement Standard

[2021]量标青证城阳字第012号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格。特发此证。

This is to certify that the measurement standard conforms with the requirements of the "Rule for the Examination of Measurement Standard" according to the Law on Metrology of the People's Republic of China.

建标单位名称 青岛斯坦德计量研究院有限公司  
Possessor of the Measurement Standard

计量标准名称 洛氏硬度计检定装置 代码 12615102  
Name of Measurement Standard Code

测量范围 (20~100) HR (A, B, C, N, T标尺)  
Measuring range

不确定度或准确度等级 (20~100) HR (A, B, C标尺)  $U = 0.5HR$ ,  $k=2$  (20~94)  
或最大允许误差  $HRN U = 0.9HR$ ,  $k=2$  (10~93)  $HRT U = 1.2HR$ ,  $k=2$   
Uncertainty/Accuracy Class/  
Maximum Permissible Error

保存地点 山东省青岛市高新区锦业路1号蓝贝智造工场D2-2区域力学  
计量室  
Installed in

发证机关(印章)  
The issuing authority  
发证日期 2021年3月26日  
有效期至 2026年3月25日  
Date issued  
Date of expiry

0010598

计量标准器 Measurement Standard	计量器具名称 Measuring Instrument	型号 Model/Type	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	制造厂及 出厂编号 Manufacturer and Series Number
主要 配套 设备 Main Auxiliary Equipment	电子秒表	PS-61	(0~99999) s	MPE: +0.5s/d	深圳市追日电子 科技有限公司 37503
	影像测量仪	VMA2515	250×150mm	MPE: (3.0+L/200) μm	天准科技股份有 限公司31606
	球压头	Φ10mm	/	/	//
开展的 检定 及 校准 项目 Verification and Calibration Items	名称 Name	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	依据的计量 检定规程或计量技术规范的编号 及名称 Verification Regulation or Technical Specification and Its Code	
	表面洛氏硬度计	(20~94) HRN (10~93) HRT	(70~94) HR15N, (42~86) HR30N, (20~77) HR45N, MPE: ±2.0HRN/ (67~93) HR15TW, (29~82) HR30T, (10~72) HR45TW MPE: ±3.0HRTW	JJG112-2003《金属洛氏硬度计(A、B、C、D、E、F、G、H、K、N、T标尺)检定规程》	
	洛氏硬度计	(20~100) HR (A、B、C标尺)	(20~88) HRA, MPE: ±1.5HRA/ (20~100) HRBW, MPE: ±2.0HRBW/ (20~70) HRC, MPE: ±1.5HRC	JJG112-2003《金属洛氏硬度计(A、B、C、D、E、F、G、H、K、N、T标尺)检定规程》	

考评员: 梁亮