

计量标准考核证书

Certificate for Examination of Measurement Standard

[2021]量标青证城阳字第021号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格。特发此证。

This is to certify that the measurement standard conforms with the requirements of the "Rule for the Examination of Measurement Standard" according to the Law on Metrology of the People's Republic of China.

建标单位名称 青岛斯坦德计量研究院有限公司
Possessor of the Measurement Standard

计量标准名称 秒表检定装置 代码 25910000
Name of Measurement Standard Code

测量范围 T: 1s ~ 3600s
Measuring range

不确定度或准确度等级
或最大允许误差 频率准确度: $U=2 \times 10^{-9}$
Uncertainty/Accuracy Class/
Maximum Permissible Error

保存地点 山东省青岛市高新区锦业路1号创业楼D2-210
Installed in

发证机关(印章)
The issuing authority
发证日期 2021年3月26日
Date issued
有效期至 2026年3月25日
Date of expiry

0010599

计量标准器 Measurement Standard	计量器具名称 Measuring Instrument	型号 Model/Type	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	制造厂及 出厂编号 Manufacturer and Series Number
	秒表检定仪	SF2002	(0.01 ~ 99999) s	频率准确度: $U=2 \times 10^{-9}$	石家庄数英仪器有限公司 3F5007003
主要配套设备 Main Auxiliary Equipment					
开展的检定及校准项目 Verification and Calibration Items	名称 Name	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	依据的计量 检定规程或计量技术规范的编号 及名称 Verification Regulation or Technical Specification and Its Code	
	电子秒表	1s~3600s	合格(MPE: 10s: $\pm 0.05s$ 1min: $\pm 0.05s$ 1h: $\pm 0.10s$)	JJG237-2010《秒表检定规程》	
	机械秒表	1s ~ 3600s	MPE: 优等及以下		

考评员: 郁黎